

CERTIFICATE OF COMPLIANCE

HAZARDOUS (CLASSIFIED) LOCATION ELECTRICAL EQUIPMENT

This certificate is issued for the following equipment:

CEL-35a/IS dBadge Noise Dosemeter
IS / I / ABCD / T4

a = 0 – 9, Firmware

Special conditions of use

- 1) *The batteries shall only be re-charged using the Casella CEL-6362 charger unit with nominal output parameters of 5.2V 220mA.*
- 2) *Parts of the enclosure are non-conducting and may generate an ignition capable level of electrostatic charge under extreme conditions. The user should ensure that the equipment is not installed or used in a location where it may be subjected to external conditions (such as high pressure steam), which might cause a build-up of electrostatic charge on non-conducting surfaces. Additionally, cleaning of the equipment should be done only with a damp cloth.*

Equipment Ratings:

Intrinsically Safe for use in Class I, Division 1, Groups, A, B, C and D Hazardous (Classified) Locations with a temperature classification of T4.

FM Approved for:

Casella CEL Limited
Bedford, MK42 7JY
United Kingdom

This certifies that the equipment described has been found to comply with the following Approval Standards and other documents:

Class 3600	1998
Class 3610	2007
Class 3810	2005

Original Project ID: 3029871

Approval Granted: *1 OCTOBER 2007*

Subsequent Revision Reports / Date Approval Amended

Report Number	Date	Report Number	Date
---------------	------	---------------	------

FM Approvals LLC



Roger L. Allard
Assistant Vice President

1 OCTOBER 2007
Date